

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
 Data File : VX038966.D
 Acq On : 20 Nov 2023 18:09
 Operator : JC/MD
 Sample : 05431-10 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22BR-20231113

Quant Time: Nov 21 00:24:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	215644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225165	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85557	50.564	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.120%
35) Dibromofluoromethane	5.385	113	70176	45.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.120%
50) Toluene-d8	8.647	98	273226	45.852	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.700%#
62) 4-Bromofluorobenzene	11.079	95	122013	51.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.300%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.794	84	1222	0.753	ug/l	# 84
27) cis-1,2-Dichloroethene	4.495	96	1915	1.138	ug/l	94
44) Trichloroethene	7.129	130	7685	4.782	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
Data File : VX038966.D
Acq On : 20 Nov 2023 18:09
Operator : JC/MD
Sample : 05431-10 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22BR-20231113

Quant Time: Nov 21 00:24:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

