

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
 Data File : VX038969.D
 Acq On : 20 Nov 2023 19:17
 Operator : JC/MD
 Sample : 05431-13 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20BR-20231113

Quant Time: Nov 21 00:25:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	214606	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	108842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84996	50.558	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.120%
35) Dibromofluoromethane	5.385	113	69825	45.050	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.100%
50) Toluene-d8	8.647	98	269094	45.377	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.760%#
62) 4-Bromofluorobenzene	11.079	95	116833	49.216	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.440%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	11362	7.423	ug/l	95
20) Methylene Chloride	2.782	84	1643	1.018	ug/l	84
27) cis-1,2-Dichloroethene	4.489	96	67955	40.631	ug/l	88
44) Trichloroethene	7.123	130	14690	9.185	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
Data File : VX038969.D
Acq On : 20 Nov 2023 19:17
Operator : JC/MD
Sample : 05431-13 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20BR-20231113

Quant Time: Nov 21 00:25:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

