

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
 Data File : VX038970.D
 Acq On : 20 Nov 2023 19:40
 Operator : JC/MD
 Sample : 05431-14 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-01-20231113

Quant Time: Nov 21 00:26:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208846	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105855	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84294	51.480	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.960%
35) Dibromofluoromethane	5.385	113	67723	44.899	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.800%
50) Toluene-d8	8.647	98	269851	46.760	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.520%
62) 4-Bromofluorobenzene	11.079	95	114126	49.402	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.800%
Target Compounds						
20) Methylene Chloride	2.788	84	3063	1.949	ug/l #	81
27) cis-1,2-Dichloroethene	4.489	96	45076	27.671	ug/l	87
44) Trichloroethene	7.123	130	59942	38.515	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112023\
Data File : VX038970.D
Acq On : 20 Nov 2023 19:40
Operator : JC/MD
Sample : 05431-14 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-01-20231113

Quant Time: Nov 21 00:26:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

