

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
 Data File : VX048627.D
 Acq On : 20 Nov 2025 11:24
 Operator : JC/MD
 Sample : Q3652-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: Nov 20 11:42:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	141582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	274881	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	301453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	157279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	103238	52.330	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.660%	
35) Dibromofluoromethane	5.379	113	87414	42.016	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	84.040%	
50) Toluene-d8	8.635	98	317797	48.977	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	97.960%	
62) 4-Bromofluorobenzene	11.061	95	137048	56.676	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	113.360%	
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	71441	76.502	ug/l	97
17) Carbon Disulfide	2.520	76	1408	0.304	ug/l	99
18) Methyl Acetate	2.697	43	2910	1.134	ug/l	95
20) Methylene Chloride	2.788	84	10951	5.704	ug/l	98
25) 2-Butanone	4.556	43	12288	8.895	ug/l #	89
43) Isopropyl Acetate	6.330	43	13748	2.494	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
Data File : VX048627.D
Acq On : 20 Nov 2025 11:24
Operator : JC/MD
Sample : Q3652-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1

Quant Time: Nov 20 11:42:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

