

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
 Data File : VX048628.D
 Acq On : 20 Nov 2025 11:45
 Operator : JC/MD
 Sample : Q3659-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TANK-1-COMP-1

Quant Time: Nov 20 12:02:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	195711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	393250	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	428037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	220456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	154803	56.765	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	113.540%	
35) Dibromofluoromethane	5.373	113	129664	43.565	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	87.120%	
50) Toluene-d8	8.635	98	476624	51.345	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	102.680%	
62) 4-Bromofluorobenzene	11.061	95	200228	57.880	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	115.760%	
Target Compounds						Qvalue
16) Acetone	2.374	43	79416	61.521	ug/l	99
20) Methylene Chloride	2.794	84	13519	5.094	ug/l	95
25) 2-Butanone	4.544	43	26464	13.858	ug/l	92
43) Isopropyl Acetate	6.330	43	14493	1.838	ug/l	94
52) Toluene	8.702	92	28937	4.449	ug/l	99
67) Ethyl Benzene	10.177	91	115540	6.772	ug/l	99
68) m/p-Xylenes	10.281	106	240667	37.456	ug/l	97
69) o-Xylene	10.622	106	242150	38.519	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
Data File : VX048628.D
Acq On : 20 Nov 2025 11:45
Operator : JC/MD
Sample : Q3659-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TANK-1-COMP-1

Quant Time: Nov 20 12:02:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

