

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
 Data File : VX048635.D
 Acq On : 20 Nov 2025 14:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Nov 21 00:33:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	142249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	288040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	314669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	170056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	126609	63.875	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 127.760%#		
35) Dibromofluoromethane	5.373	113	106562	48.880	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.760%		
50) Toluene-d8	8.634	98	363316	53.435	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 106.860%		
62) 4-Bromofluorobenzene	11.061	95	162620	64.179	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 128.360%#		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.940	59	3574	9.516	ug/l	96
16) Acetone	2.379	43	2036	2.170	ug/l #	80
95) Naphthalene	13.762	128	3066	0.250	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
Data File : VX048635.D
Acq On : 20 Nov 2025 14:09
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Nov 21 00:33:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

