

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
 Data File : VX048637.D
 Acq On : 20 Nov 2025 15:15
 Operator : JC/MD
 Sample : Q3685-01 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 4197

Quant Time: Nov 21 00:33:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	196012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	396369	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	414483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	221674	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	150052	54.938	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.880%
35) Dibromofluoromethane	5.373	113	127877	42.626	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.260%
50) Toluene-d8	8.635	98	459872	49.150	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.061	95	195739	56.137	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.280%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	10130	7.835	ug/l	92
18) Methyl Acetate	2.703	43	4349	1.224	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112025\
Data File : VX048637.D
Acq On : 20 Nov 2025 15:15
Operator : JC/MD
Sample : Q3685-01 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
4197

Quant Time: Nov 21 00:33:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

