

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112118\
 Data File : VX006125.D
 Acq On : 21 Nov 2018 15:40
 Operator : JC/MD
 Sample : J6016-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-112018

Quant Time: Nov 22 03:07:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	351659	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	521800	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	479493	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	236620	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	216326	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
35) Dibromofluoromethane	5.50	113	172664	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
50) Toluene-d8	8.71	98	631387	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	11.14	95	214960	44.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.68%	
Target Compounds						
17) Carbon Disulfide	2.57	76	30350	4.427	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112118\
Data File : VX006125.D
Acq On : 21 Nov 2018 15:40
Operator : JC/MD
Sample : J6016-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-112018

Quant Time: Nov 22 03:07:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

