

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112118\
 Data File : VX006127.D
 Acq On : 21 Nov 2018 16:59
 Operator : JC/MD
 Sample : J5981-05ME
 Misc : 2.82g/10ml/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DTW-03(8-10)ME

Quant Time: Nov 26 10:43:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	352155	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.86	114	530035	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	496465	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	274192	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	212487	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
35) Dibromofluoromethane	5.51	113	165784	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
50) Toluene-d8	8.71	98	653138	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	11.14	95	243048	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
Target Compounds						
					Qvalue	
16) Acetone	2.49	43	5989	3.307	ug/l	94
17) Carbon Disulfide	2.53	76	1416171	206.258	ug/l	98
52) Toluene	8.78	92	12138	1.381	ug/l	96
67) Ethyl Benzene	10.25	91	65604	3.934	ug/l	96
68) m/p-Xylenes	10.36	106	41381	6.373	ug/l	99
69) o-Xylene	10.70	106	22604	3.645	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	31968	2.241	ug/l	99
95) Naphthalene	13.83	128	3214993	177.157	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112118\
Data File : VX006127.D
Acq On : 21 Nov 2018 16:59
Operator : JC/MD
Sample : J5981-05ME
Misc : 2.82g/10ml/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DTW-03(8-10)ME

Quant Time: Nov 26 10:43:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

