

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112118\
 Data File : VX006128.D
 Acq On : 21 Nov 2018 17:26
 Operator : JC/MD
 Sample : J6031-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 22 03:16:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	376645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	566817	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	518908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	266700	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	227994	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
35) Dibromofluoromethane	5.50	113	179844	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
50) Toluene-d8	8.71	98	689627	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
62) 4-Bromofluorobenzene	11.13	95	243114	46.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.38%	
Target Compounds						
17) Carbon Disulfide	2.57	76	14488	1.973	ug/l #	94
95) Naphthalene	13.83	128	63847	3.617	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112118\
Data File : VX006128.D
Acq On : 21 Nov 2018 17:26
Operator : JC/MD
Sample : J6031-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 22 03:16:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

