

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112118\
 Data File : VX006145.D
 Acq On : 22 Nov 2018 01:51
 Operator : JC/MD
 Sample : J6047-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP2-MW-102

Quant Time: Nov 22 04:19:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	333243	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	504850	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	461659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	225894	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	205939	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
35) Dibromofluoromethane	5.50	113	166571	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
50) Toluene-d8	8.71	98	603516	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	11.14	95	205349	44.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.56%	
Target Compounds						
64) Tetrachloroethene	9.33	164	3204	0.843	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112118\
Data File : VX006145.D
Acq On : 22 Nov 2018 01:51
Operator : JC/MD
Sample : J6047-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP2-MW-102

Quant Time: Nov 22 04:19:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

