

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112118\
 Data File : VX006150.D
 Acq On : 22 Nov 2018 04:04
 Operator : JC/MD
 Sample : J6032-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MMTW-DUP

Quant Time: Nov 22 04:31:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	287233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	424859	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	421135	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182794	54.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.84%	
35) Dibromofluoromethane	5.50	113	145859	53.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.96%	
50) Toluene-d8	8.71	98	509067	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
62) 4-Bromofluorobenzene	11.14	95	186496	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.56%	
Target Compounds						
16) Acetone	2.44	43	3663	2.480	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112118\
Data File : VX006150.D
Acq On : 22 Nov 2018 04:04
Operator : JC/MD
Sample : J6032-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MMTW-DUP

Quant Time: Nov 22 04:31:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

