

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112118\
 Data File : VX006151.D
 Acq On : 22 Nov 2018 04:30
 Operator : JC/MD
 Sample : J6031-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VTW-01

Quant Time: Nov 22 06:35:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	326064	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	489373	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	451413	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	221706	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	204449	54.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.22%	
35) Dibromofluoromethane	5.50	113	164301	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
50) Toluene-d8	8.71	98	587667	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
62) 4-Bromofluorobenzene	11.13	95	200794	44.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.32%	
Target Compounds						
11) Tert butyl alcohol	3.06	59	5937	7.387	ug/l #	82
16) Acetone	2.44	43	2266	1.352	ug/l	95
19) Methyl tert-butyl Ether	3.20	73	48366	5.943	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112118\
Data File : VX006151.D
Acq On : 22 Nov 2018 04:30
Operator : JC/MD
Sample : J6031-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VTW-01

Quant Time: Nov 22 06:35:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

