

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112118\
 Data File : VX006152.D
 Acq On : 22 Nov 2018 04:57
 Operator : JC/MD
 Sample : J6031-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VTW-02

Quant Time: Nov 22 06:36:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	322047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	479898	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	446136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	219585	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	200002	53.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.20%	
35) Dibromofluoromethane	5.50	113	161189	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
50) Toluene-d8	8.71	98	580882	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.14	95	196777	44.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.28%	
Target Compounds						
16) Acetone	2.44	43	3498	2.112	ug/l	95
19) Methyl tert-butyl Ether	3.20	73	9980	1.242	ug/l	94
95) Naphthalene	13.83	128	5833	0.401	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112118\
Data File : VX006152.D
Acq On : 22 Nov 2018 04:57
Operator : JC/MD
Sample : J6031-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VTW-02

Quant Time: Nov 22 06:36:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

