


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112118\
Data File : VX006137.D
Acq On : 21 Nov 2018 21:52
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Title : SW846 8260
Last Update : Wed Nov 21 08:15:54 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	7013	PASS
75	95	30	60	49.1	18441	PASS
95	95	100	100	100.0	37568	PASS
96	95	5	9	6.8	2550	PASS
173	174	0.00	2	1.2	455	PASS
174	95	50	100	97.0	36437	PASS
175	174	5	9	7.7	2808	PASS
176	174	95	101	96.4	35117	PASS
177	176	5	9	6.8	2388	PASS