

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112119\
 Data File : VX013496.D
 Acq On : 21 Nov 2019 12:30
 Operator : JC/SP
 Sample : K5958-03DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3DL

Quant Time: Nov 22 07:55:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	627324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	997188	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	900978	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	467314	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	365372	49.16	ug/l	0.00
Spiked Amount			Recovery	=	98.32%	
35) Dibromofluoromethane	5.49	113	301628	48.23	ug/l	0.00
Spiked Amount			Recovery	=	96.46%	
50) Toluene-d8	8.71	98	1204943	51.10	ug/l	0.00
Spiked Amount			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	11.14	95	436987	51.35	ug/l	0.00
Spiked Amount			Recovery	=	102.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	7.45	83	20722	1.944	ug/l	97
67) Ethyl Benzene	10.25	91	999982	31.764	ug/l	98
68) m/p-Xylenes	10.35	106	336448	28.259	ug/l	99
69) o-Xylene	10.70	106	246583	21.154	ug/l	99
73) Isopropylbenzene	11.01	105	1213699	39.172	ug/l	100
78) n-propylbenzene	11.35	91	2445930	70.868	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	804724	31.182	ug/l	100
83) tert-Butylbenzene	11.76	119	120954	4.656	ug/l	88
84) 1,2,4-Trimethylbenzene	11.81	105	9179669	354.580	ug/l	94
85) sec-Butylbenzene	11.94	105	536101	17.955	ug/l #	75
86) p-Isopropyltoluene	12.06	119	250546	9.113	ug/l	99
89) n-Butylbenzene	12.39	91	340144	14.178	ug/l #	67
95) Naphthalene	13.83	128	323095	10.031	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112119\
Data File : VX013496.D
Acq On : 21 Nov 2019 12:30
Operator : JC/SP
Sample : K5958-03DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3DL

Quant Time: Nov 22 07:55:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

