

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112119\
 Data File : VX013500.D
 Acq On : 21 Nov 2019 14:03
 Operator : JC/SP
 Sample : K5967-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WASTE-DRUM

Quant Time: Nov 22 08:23:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	660336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1064994	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	950359	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	442758	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	373477	47.74	ug/l	0.01
Spiked Amount			Recovery	=	95.48%	
35) Dibromofluoromethane	5.51	113	269128	40.29	ug/l	0.02
Spiked Amount			Recovery	=	80.58%	
50) Toluene-d8	8.71	98	1280773	50.86	ug/l	0.00
Spiked Amount			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	11.14	95	427577	47.05	ug/l	0.00
Spiked Amount			Recovery	=	94.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.32	50	93945	11.386	ug/l	99
16) Acetone	2.52	43	687387	176.162	ug/l	99
18) Methyl Acetate	2.81	43	19014542	1765.080	ug/l	99
29) Tetrahydrofuran	5.19	42	248419	70.422	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	11754	0.479	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112119\
Data File : VX013500.D
Acq On : 21 Nov 2019 14:03
Operator : JC/SP
Sample : K5967-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WASTE-DRUM

Quant Time: Nov 22 08:23:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

