

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112119\
 Data File : VX013508.D
 Acq On : 21 Nov 2019 17:10
 Operator : JC/SP
 Sample : K5958-03DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3DL

Quant Time: Nov 22 07:59:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	654120	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1052538	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	960321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	505738	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	382632	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	5.49	113	317811	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
50) Toluene-d8	8.71	98	1277316	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.14	95	469794	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
Target Compounds						
					Qvalue	
67) Ethyl Benzene	10.25	91	269238	8.024	ug/l	99
68) m/p-Xylenes	10.35	106	90239	7.111	ug/l	99
69) o-Xylene	10.70	106	66038	5.315	ug/l	98
73) Isopropylbenzene	11.01	105	327838	9.777	ug/l	99
78) n-propylbenzene	11.35	91	673302	18.026	ug/l	100
80) 1,3,5-Trimethylbenzene	11.50	105	234841	8.408	ug/l	99
83) tert-Butylbenzene	11.76	119	33446	1.190	ug/l	85
84) 1,2,4-Trimethylbenzene	11.81	105	2766674	98.748	ug/l	97
85) sec-Butylbenzene	11.94	105	160483	4.967	ug/l #	75
86) p-Isopropyltoluene	12.06	119	73277	2.463	ug/l	99
89) n-Butylbenzene	12.38	91	100479	3.870	ug/l #	67
95) Naphthalene	13.83	128	87600	2.513	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112119\
Data File : VX013508.D
Acq On : 21 Nov 2019 17:10
Operator : JC/SP
Sample : K5958-03DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3DL

Quant Time: Nov 22 07:59:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

