

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112119\
 Data File : VX013511.D
 Acq On : 21 Nov 2019 18:20
 Operator : JC/SP
 Sample : K5962-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22BR-20191120

Quant Time: Nov 22 08:46:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	706860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1140697	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1025870	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	453711	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	414537	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	5.49	113	344289	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
50) Toluene-d8	8.71	98	1372931	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
62) 4-Bromofluorobenzene	11.14	95	454754	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	144774	18.306	ug/l	90
44) Trichloroethene	7.21	130	122831	14.639	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112119\
Data File : VX013511.D
Acq On : 21 Nov 2019 18:20
Operator : JC/SP
Sample : K5962-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22BR-20191120

Quant Time: Nov 22 08:46:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

