

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019495.D
 Acq On : 20 Nov 2020 14:49
 Operator : JC/MD
 Sample : VX1121WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1121WBL01

Quant Time: Nov 21 04:01:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	300846	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.82	114	484077	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	432050	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	192044	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	182482	46.78	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.56%	
35) Dibromofluoromethane		5.46	113	143978	47.38	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.76%	
50) Toluene-d8		8.70	98	580272	49.14	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.28%	
62) 4-Bromofluorobenzene		11.12	95	199648	47.00	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019495.D
Acq On : 20 Nov 2020 14:49
Operator : JC/MD
Sample : VX1121WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1121WBL01

Quant Time: Nov 21 04:01:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

