

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019500.D
 Acq On : 20 Nov 2020 17:02
 Operator : JC/MD
 Sample : L4837-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 21 04:14:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	287994	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	463409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	401285	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	174047	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	176587	47.29	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.58%
35) Dibromofluoromethane	5.46	113	139160	47.84	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.68%
50) Toluene-d8	8.70	98	552862	48.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.82%
62) 4-Bromofluorobenzene	11.12	95	180571	44.40	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019500.D
Acq On : 20 Nov 2020 17:02
Operator : JC/MD
Sample : L4837-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 21 04:14:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

