

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019501.D
 Acq On : 20 Nov 2020 17:26
 Operator : JC/MD
 Sample : L4837-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Nov 21 04:15:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	273880	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	447624	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	381019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	157675	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	168964	47.58	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.16%
35) Dibromofluoromethane	5.46	113	133430	47.49	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.98%
50) Toluene-d8	8.70	98	524885	48.07	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.14%
62) 4-Bromofluorobenzene	11.12	95	168299	42.84	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019501.D
Acq On : 20 Nov 2020 17:26
Operator : JC/MD
Sample : L4837-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Nov 21 04:15:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

