

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019502.D
 Acq On : 20 Nov 2020 17:49
 Operator : JC/MD
 Sample : L4827-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 201118001-05-TRIP-BLANK

Quant Time: Nov 21 04:20:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	272113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	438933	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	368136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	139937	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	167169	47.38	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.76%		
35) Dibromofluoromethane	5.46	113	130750	47.46	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.92%		
50) Toluene-d8	8.70	98	517765	48.36	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.72%		
62) 4-Bromofluorobenzene	11.12	95	160279	41.61	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	83.22%		

Target Compounds					Qvalue
16) Acetone	2.43	43	3908	2.864 ug/l	94
20) Methylene Chloride	2.82	84	2427	0.712 ug/l	89
30) Chloroform	5.17	83	6321	1.129 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019502.D
Acq On : 20 Nov 2020 17:49
Operator : JC/MD
Sample : L4827-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
201118001-05-TRIP-BLANK

Quant Time: Nov 21 04:20:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

