

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019504.D
 Acq On : 20 Nov 2020 18:36
 Operator : JC/MD
 Sample : L4836-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-02

Quant Time: Nov 21 04:32:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	272709	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	448220	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	401203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	167277	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	169690	47.99	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.98%
35) Dibromofluoromethane	5.46	113	134318	47.74	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.48%
50) Toluene-d8	8.70	98	536837	49.10	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.20%
62) 4-Bromofluorobenzene	11.12	95	181194	46.06	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.12%

Target Compounds

					Qvalue	
8) Diethyl Ether	2.17	74	1872	0.989	ug/l	98
11) Tert butyl alcohol	3.04	59	4099	6.125	ug/l #	82
16) Acetone	2.43	43	3774	2.760	ug/l	98
18) Methyl Acetate	2.75	43	1168	0.978	ug/l	99
29) Tetrahydrofuran	5.11	42	1638	1.250	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019504.D
Acq On : 20 Nov 2020 18:36
Operator : JC/MD
Sample : L4836-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-02

Quant Time: Nov 21 04:32:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

