

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019508.D
 Acq On : 20 Nov 2020 20:09
 Operator : JC/MD
 Sample : L4772-02DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4DL

Quant Time: Nov 21 04:47:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	250495	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	408544	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	356520	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	161258	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	153942	47.40	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.80%		
35) Dibromofluoromethane	5.46	113	120218	46.88	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.76%		
50) Toluene-d8	8.70	98	479949	48.16	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.32%		
62) 4-Bromofluorobenzene	11.12	95	167836	46.81	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.62%		

Target Compounds

						Qvalue
27) cis-1,2-Dichloroethene	4.56	96	1834	0.567	ug/l	68
67) Ethyl Benzene	10.24	91	70532	5.493	ug/l	97
68) m/p-Xylenes	10.34	106	17473	3.652	ug/l	98
69) o-Xylene	10.68	106	1954	0.427	ug/l	94
73) Isopropylbenzene	11.00	105	83852	7.517	ug/l	99
78) n-propylbenzene	11.34	91	151187	11.814	ug/l	98
80) 1,3,5-Trimethylbenzene	11.49	105	28325	3.027	ug/l	100
83) tert-Butylbenzene	11.76	119	11752	1.298	ug/l	88
84) 1,2,4-Trimethylbenzene	11.79	105	956236	100.990	ug/l	99
85) sec-Butylbenzene	11.93	105	49637	4.609	ug/l	79
86) p-Isopropyltoluene	12.05	119	14750	1.508	ug/l	95
89) n-Butylbenzene	12.37	91	25667	2.903	ug/l #	70
95) Naphthalene	13.82	128	51821	6.345	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019508.D
Acq On : 20 Nov 2020 20:09
Operator : JC/MD
Sample : L4772-02DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4DL

Quant Time: Nov 21 04:47:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

