

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019516.D
 Acq On : 21 Nov 2020 00:50
 Operator : JC/MD
 Sample : PB133122ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133122ZHE#01

Quant Time: Nov 21 05:58:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	276173	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	454151	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	397869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	165009	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	171485	47.89	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.78%		
35) Dibromofluoromethane	5.46	113	138525	48.59	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.18%		
50) Toluene-d8	8.70	98	537115	48.49	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.98%		
62) 4-Bromofluorobenzene	11.12	95	176699	44.33	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.66%		

Target Compounds

					Qvalue
16) Acetone	2.43	43	15030	10.854	ug/l 99
18) Methyl Acetate	2.76	43	3261	1.583	ug/l 90
20) Methylene Chloride	2.83	84	8732	2.523	ug/l 94
43) Isopropyl Acetate	6.42	43	32121	4.781	ug/l 99
95) Naphthalene	13.82	128	985	2.255	ug/l # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019516.D
Acq On : 21 Nov 2020 00:50
Operator : JC/MD
Sample : PB133122ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133122ZHE#01

Quant Time: Nov 21 05:58:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

