

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019523.D
 Acq On : 21 Nov 2020 03:34
 Operator : JC/MD
 Sample : L4779-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-1654-1658

Quant Time: Nov 21 06:52:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	303583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	504336	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	436586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199815	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	183781	46.69	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.38%		
35) Dibromofluoromethane	5.46	113	147916	46.72	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.44%		
50) Toluene-d8	8.70	98	593764	48.27	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.54%		
62) 4-Bromofluorobenzene	11.12	95	212840	48.09	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.18%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.04	59	87614	117.614	ug/l	99
16) Acetone	2.43	43	375213	246.493	ug/l	99
25) 2-Butanone	4.64	43	165549	72.398	ug/l	100
30) Chloroform	5.18	83	1621	0.259	ug/l	89
40) Benzene	6.11	78	13506	0.930	ug/l	100
51) 4-Methyl-2-Pentanone	8.62	43	32966	7.577	ug/l	98
52) Toluene	8.76	92	15891	1.788	ug/l	97
59) 2-Hexanone	9.48	43	74205	22.487	ug/l	100
67) Ethyl Benzene	10.23	91	8798	0.559	ug/l	98
68) m/p-Xylenes	10.34	106	7987	1.363	ug/l	100
69) o-Xylene	10.68	106	5884	1.051	ug/l	98
78) n-propylbenzene	11.34	91	3856	0.243	ug/l	94
80) 1,3,5-Trimethylbenzene	11.49	105	4797	0.414	ug/l	95
84) 1,2,4-Trimethylbenzene	11.79	105	25692	2.190	ug/l	96
89) n-Butylbenzene	12.37	91	3906	0.357	ug/l #	40
95) Naphthalene	13.82	128	126565	10.391	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019523.D
Acq On : 21 Nov 2020 03:34
Operator : JC/MD
Sample : L4779-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-1654-1658

Quant Time: Nov 21 06:52:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

