

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019528.D
 Acq On : 21 Nov 2020 05:31
 Operator : JC/MD
 Sample : L4826-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB-2020-WW-000-04

Quant Time: Nov 21 08:00:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	342401	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	551949	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	476759	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	219315	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	201302	45.34	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.68%
35) Dibromofluoromethane	5.46	113	163637	47.23	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.46%
50) Toluene-d8	8.70	98	650368	48.31	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.62%
62) 4-Bromofluorobenzene	11.12	95	232625	48.03	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.06%

Target Compounds

					Qvalue
16) Acetone	2.43	43	18724	10.906	ug/l 94
17) Carbon Disulfide	2.54	76	3366	0.369	ug/l 98
20) Methylene Chloride	2.84	84	1502	0.350	ug/l # 78
27) cis-1,2-Dichloroethene	4.56	96	1795	0.406	ug/l # 21
40) Benzene	6.11	78	10826	0.681	ug/l 92
44) Trichloroethene	7.18	130	49130	11.797	ug/l 94
64) Tetrachloroethene	9.32	164	1019137	271.946	ug/l 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019528.D
Acq On : 21 Nov 2020 05:31
Operator : JC/MD
Sample : L4826-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB-2020-WW-000-04

Quant Time: Nov 21 08:00:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

