

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019537.D
 Acq On : 21 Nov 2020 09:24
 Operator : JC/MD
 Sample : L4687-04ME
 Misc : 5.78g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS-3ME

Quant Time: Nov 23 00:27:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	301687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	485360	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	422358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	190897	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	178294	45.58	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.16%
35) Dibromofluoromethane	5.47	113	139968	45.94	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.88%
50) Toluene-d8	8.70	98	568394	48.01	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.02%
62) 4-Bromofluorobenzene	11.12	95	200873	47.16	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.32%

Target Compounds

					Qvalue
44) Trichloroethene	7.18	130	4247	1.160 ug/l	97
64) Tetrachloroethene	9.32	164	333319	100.399 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019537.D
Acq On : 21 Nov 2020 09:24
Operator : JC/MD
Sample : L4687-04ME
Misc : 5.78g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS-3ME

Quant Time: Nov 23 00:27:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

