

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112120\
 Data File : VX019538.D
 Acq On : 21 Nov 2020 09:47
 Operator : JC/MD
 Sample : L4687-06ME
 Misc : 5.77g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS-5ME

Quant Time: Nov 23 00:29:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	292654	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	475315	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	408672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	178804	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	173069	45.61	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.22%
35) Dibromofluoromethane	5.47	113	134837	45.19	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.38%
50) Toluene-d8	8.70	98	551662	47.58	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.16%
62) 4-Bromofluorobenzene	11.12	95	191923	46.01	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.02%

Target Compounds

					Qvalue
27) cis-1,2-Dichloroethene	4.56	96	39627	10.480	ug/l 91
44) Trichloroethene	7.18	130	31685	8.835	ug/l 99
64) Tetrachloroethene	9.32	164	92363	28.752	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112120\
Data File : VX019538.D
Acq On : 21 Nov 2020 09:47
Operator : JC/MD
Sample : L4687-06ME
Misc : 5.77g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS-5ME

Quant Time: Nov 23 00:29:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

