

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112120\
 Data File : VX019534.D
 Acq On : 21 Nov 2020 08:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 00:20:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.486	0.476	2.1	100	0.00
3 P	Chloromethane	0.558	0.521	6.6	97	0.00
4 C	Vinyl Chloride	0.601	0.598	0.5#	104	0.00
5 T	Bromomethane	0.227	0.189	16.7	87	0.00
6 T	Chloroethane	0.166	0.157	5.4	104	0.00
7 T	Trichlorofluoromethane	0.867	0.863	0.5	107	0.00
8 T	Diethyl Ether	0.347	0.371	-6.9	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.461	4.8	100	0.00
10 T	Methyl Iodide	0.635	0.582	8.3	92	0.00
11 T	Tert butyl alcohol	0.123	0.145	-17.9	126	0.00
12 CM	1,1-Dichloroethene	0.496	0.500	-0.8#	108	0.00
13 T	Acrolein	0.082	0.066	19.5	88	0.00
14 T	Allyl chloride	0.778	0.755	3.0	105	0.00
15 T	Acrylonitrile	0.289	0.302	-4.5	108	0.00
16 T	Acetone	0.251	0.242	3.6	104	0.00
17 T	Carbon Disulfide	1.334	1.278	4.2	104	0.00
18 T	Methyl Acetate	0.647	0.647	0.0	112	0.00
19 T	Methyl tert-butyl Ether	1.713	1.908	-11.4	116	0.00
20 T	Methylene Chloride	0.627	0.607	3.2	107	0.00
21 T	trans-1,2-Dichloroethene	0.562	0.564	-0.4	107	0.00
22 T	Diisopropyl ether	1.609	1.710	-6.3	110	0.00
23 T	Vinyl Acetate	1.349	1.462	-8.4	110	0.00
24 P	1,1-Dichloroethane	0.984	1.026	-4.3	111	0.00
25 T	2-Butanone	0.377	0.385	-2.1	106	0.00
26 T	2,2-Dichloropropane	0.859	0.614	28.5#	76	0.00
27 T	cis-1,2-Dichloroethene	0.646	0.684	-5.9	111	0.00
28 T	Bromochloromethane	0.470	0.448	4.7	111	0.00
29 T	Tetrahydrofuran	0.240	0.249	-3.8	109	0.00
30 C	Chloroform	1.029	1.078	-4.8#	109	0.00
31 T	Cyclohexane	0.843	0.820	2.7	103	0.00
32 T	1,1,1-Trichloroethane	0.884	0.940	-6.3	111	0.00
33 S	1,2-Dichloroethane-d4	0.648	0.661	-2.0	112	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	113	0.00
35 S	Dibromofluoromethane	0.314	0.308	1.9	113	0.00
36 T	1,1-Dichloropropene	0.469	0.438	6.6	106	0.00
37 T	Ethyl Acetate	0.440	0.439	0.2	110	0.00
38 T	Carbon Tetrachloride	0.454	0.445	2.0	109	0.00
39 T	Methylcyclohexane	0.546	0.468	14.3	95	0.00
40 TM	Benzene	1.440	1.382	4.0	108	0.00
41 T	Methacrylonitrile	0.243	0.238	2.1	109	0.00
42 TM	1,2-Dichloroethane	0.494	0.485	1.8	111	0.00
43 T	Isopropyl Acetate	0.740	0.736	0.5	113	0.00
44 TM	Trichloroethene	0.377	0.361	4.2	109	0.00
45 C	1,2-Dichloropropane	0.355	0.346	2.5#	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112120\
 Data File : VX019534.D
 Acq On : 21 Nov 2020 08:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 00:20:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.243	0.242	0.4	111	0.00
47 T	Bromodichloromethane	0.471	0.480	-1.9	110	0.00
48 T	Methyl methacrylate	0.346	0.358	-3.5	111	0.00
49 T	1,4-Dioxane	0.008	0.008	0.0	116	0.00
50 S	Toluene-d8	1.220	1.140	6.6	107	0.00
51 T	4-Methyl-2-Pentanone	0.431	0.433	-0.5	108	0.00
52 CM	Toluene	0.881	0.857	2.7#	106	0.00
53 T	t-1,3-Dichloropropene	0.509	0.499	2.0	102	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.546	3.2	103	0.00
55 T	1,1,2-Trichloroethane	0.359	0.349	2.8	105	0.00
56 T	Ethyl methacrylate	0.508	0.545	-7.3	113	0.00
57 T	1,3-Dichloropropane	0.601	0.588	2.2	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.298	0.281	5.7	106	0.00
59 T	2-Hexanone	0.327	0.326	0.3	107	0.00
60 T	Dibromochloromethane	0.351	0.364	-3.7	109	0.00
61 T	1,2-Dibromoethane	0.370	0.364	1.6	105	0.00
62 S	4-Bromofluorobenzene	0.439	0.424	3.4	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.393	0.347	11.7	96	0.00
65 PM	Chlorobenzene	1.028	1.029	-0.1	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.389	-4.0	107	0.00
67 C	Ethyl Benzene	1.801	1.832	-1.7#	103	0.00
68 T	m/p-Xylenes	0.671	0.683	-1.8	102	0.00
69 T	o-Xylene	0.641	0.667	-4.1	104	0.00
70 T	Styrene	1.065	1.130	-6.1	103	0.00
71 P	Bromoform	0.273	0.296	-8.4	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.459	3.683	-6.5	102	0.00
74 T	N-amyl acetate	1.296	1.528	-17.9	113	0.00
75 P	1,1,2,2-Tetrachloroethane	1.237	1.301	-5.2	107	0.00
76 T	1,2,3-Trichloropropane	1.091	1.148	-5.2	105	0.00
77 T	Bromobenzene	0.883	0.919	-4.1	104	0.00
78 T	n-propylbenzene	3.968	4.103	-3.4	99	0.00
79 T	2-Chlorotoluene	2.406	2.535	-5.4	103	0.00
80 T	1,3,5-Trimethylbenzene	2.902	3.111	-7.2	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.371	0.371	0.0	97	0.00
82 T	4-Chlorotoluene	2.833	2.930	-3.4	100	0.00
83 T	tert-Butylbenzene	2.807	3.004	-7.0	103	0.00
84 T	1,2,4-Trimethylbenzene	2.936	3.136	-6.8	100	0.00
85 T	sec-Butylbenzene	3.340	3.388	-1.4	96	0.00
86 T	p-Isopropyltoluene	3.032	3.102	-2.3	96	0.00
87 T	1,3-Dichlorobenzene	1.622	1.634	-0.7	100	0.00
88 T	1,4-Dichlorobenzene	1.670	1.629	2.5	99	0.00
89 T	n-Butylbenzene	2.741	2.580	5.9	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112120\
 Data File : VX019534.D
 Acq On : 21 Nov 2020 08:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 00:20:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.507	0.540	-6.5	102	0.00
91 T	1,2-Dichlorobenzene	1.629	1.649	-1.2	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.252	0.285	-13.1	111	0.00
93 T	1,2,4-Trichlorobenzene	1.046	1.067	-2.0	100	0.00
94 T	Hexachlorobutadiene	0.476	0.403	15.3	87	0.00
95 T	Naphthalene	3.237	3.774	-16.6	106	0.00
96 T	1,2,3-Trichlorobenzene	1.046	1.088	-4.0	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6