


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112120\
 Data File : VX019510.D
 Acq On : 20 Nov 2020 21:43
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Title : SW846 8260
 Last Update : Sat Nov 21 01:57:45 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	11621	PASS
75	95	30	60	49.0	32018	PASS
95	95	100	100	100.0	65301	PASS
96	95	5	9	6.9	4478	PASS
173	174	0.00	2	0.6	285	PASS
174	95	50	100	78.0	50944	PASS
175	174	5	9	6.9	3514	PASS
176	174	95	101	98.2	50005	PASS
177	176	5	9	7.0	3482	PASS