

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112122\
 Data File : VX032945.D
 Acq On : 21 Nov 2022 12:27
 Operator : JC/MD
 Sample : N5672-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20221116DL

Quant Time: Nov 22 05:11:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	181956	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78872	52.006	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.020%	
35) Dibromofluoromethane	5.385	113	63288	48.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.520%	
50) Toluene-d8	8.647	98	228405	50.609	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.220%	
62) 4-Bromofluorobenzene	11.079	95	86579	49.111	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.220%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	2760	2.355	ug/l	97
44) Trichloroethene	7.129	130	75116	52.171	ug/l	98
64) Tetrachloroethene	9.281	164	750	0.629	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112122\
Data File : VX032945.D
Acq On : 21 Nov 2022 12:27
Operator : JC/MD
Sample : N5672-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20221116DL

Quant Time: Nov 22 05:11:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

