

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112122\
 Data File : VX032952.D
 Acq On : 21 Nov 2022 15:09
 Operator : JC/MD
 Sample : N5672-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20221116

Quant Time: Nov 22 05:12:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	167033	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72650	52.000	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.000%
35) Dibromofluoromethane	5.385	113	58507	49.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.220%
50) Toluene-d8	8.647	98	210261	50.751	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.500%
62) 4-Bromofluorobenzene	11.079	95	78218	48.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.660%
Target Compounds						
44) Trichloroethene	7.129	130	9417	7.125	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112122\
Data File : VX032952.D
Acq On : 21 Nov 2022 15:09
Operator : JC/MD
Sample : N5672-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20221116

Quant Time: Nov 22 05:12:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

