

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112122\  
 Data File : VX032958.D  
 Acq On : 21 Nov 2022 17:28  
 Operator : JC/MD  
 Sample : N5729-08  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TT161S-20221117

Quant Time: Nov 22 05:12:55 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X102622W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 27 08:28:06 2022  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                       |                |      |                     |        |        |          |
| Internal Standards          |                |      |                     |        |        |          |
| 1) Pentafluorobenzene       | 5.550          | 168  | 99596               | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763          | 114  | 166364              | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055         | 117  | 149379              | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024         | 152  | 67914               | 50.000 | ug/l   | 0.00     |
|                             |                |      |                     |        |        |          |
| System Monitoring Compounds |                |      |                     |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958          | 65   | 71169               | 51.428 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 78 - 117 |      | Recovery = 102.860% |        |        |          |
| 35) Dibromofluoromethane    | 5.385          | 113  | 57230               | 48.227 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = 96.460%  |        |        |          |
| 50) Toluene-d8              | 8.647          | 98   | 209773              | 50.837 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 92 - 112 |      | Recovery = 101.680% |        |        |          |
| 62) 4-Bromofluorobenzene    | 11.079         | 95   | 77975               | 48.376 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 83 - 123 |      | Recovery = 96.760%  |        |        |          |
|                             |                |      |                     |        |        |          |
| Target Compounds            |                |      |                     |        |        | Qvalue   |
| 12) 1,1-Dichloroethene      | 2.319          | 96   | 658                 | 0.659  | ug/l   | 92       |
| 19) Methyl tert-butyl Ether | 3.117          | 73   | 1455                | 0.414  | ug/l   | 92       |
| 20) Methylene Chloride      | 2.794          | 84   | 382                 | 0.331  | ug/l # | 72       |
| 24) 1,1-Dichloroethane      | 3.611          | 63   | 6051                | 3.055  | ug/l   | 97       |
| 44) Trichloroethene         | 7.135          | 130  | 2900                | 2.203  | ug/l   | 89       |
| 64) Tetrachloroethene       | 9.269          | 164  | 392                 | 0.359  | ug/l # | 72       |
| -----                       |                |      |                     |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112122\  
Data File : VX032958.D  
Acq On : 21 Nov 2022 17:28  
Operator : JC/MD  
Sample : N5729-08  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TT161S-20221117

Quant Time: Nov 22 05:12:55 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X102622W.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 27 08:28:06 2022  
Response via : Initial Calibration

