

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
 Data File : VX038981.D
 Acq On : 21 Nov 2023 13:55
 Operator : JC/MD
 Sample : 05431-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18BR-20231113DL

Quant Time: Nov 22 03:19:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	101170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	207493	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	109207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84180	51.638	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.280%
35) Dibromofluoromethane	5.385	113	67461	45.017	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.040%
50) Toluene-d8	8.647	98	266778	46.529	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.060%
62) 4-Bromofluorobenzene	11.079	95	116702	50.846	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.700%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.319	96	533	0.423	ug/l	94
20) Methylene Chloride	2.788	84	2051	1.311	ug/l #	89
27) cis-1,2-Dichloroethene	4.501	96	3336	2.057	ug/l	80
44) Trichloroethene	7.123	130	92747	59.982	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
Data File : VX038981.D
Acq On : 21 Nov 2023 13:55
Operator : JC/MD
Sample : 05431-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18BR-20231113DL

Quant Time: Nov 22 03:19:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

