

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
 Data File : VX038982.D
 Acq On : 21 Nov 2023 14:18
 Operator : JC/MD
 Sample : 05431-06DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 45BR-20231113DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/22/2023
 Supervised By : Mahesh Dadoda 11/22/2023

Quant Time: Nov 22 03:20:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	99783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	208687	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	106586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82439	51.273	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.540%			
35) Dibromofluoromethane	5.385	113	66656	44.225	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 88.460%			
50) Toluene-d8	8.647	98	262885	45.588	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 91.180%#			
62) 4-Bromofluorobenzene	11.079	95	112166	48.590	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.180%			
Target Compounds						Qvalue
20) Methylene Chloride	2.788	84	2172	1.408	ug/l	90
27) cis-1,2-Dichloroethene	4.483	96	12931	8.084	ug/l	84
44) Trichloroethene	7.123	130	51085	32.849	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
Data File : VX038982.D
Acq On : 21 Nov 2023 14:18
Operator : JC/MD
Sample : 05431-06DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
45BR-20231113DL

Quant Time: Nov 22 03:20:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/22/2023
Supervised By :Mahesh Dadoda 11/22/2023

