

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
 Data File : VX038983.D
 Acq On : 21 Nov 2023 14:41
 Operator : JC/MD
 Sample : 05431-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-01-20231113

Quant Time: Nov 22 03:20:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	101126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208230	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82963	50.913	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.820%
35) Dibromofluoromethane	5.385	113	66598	44.284	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.560%
50) Toluene-d8	8.647	98	259087	45.028	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.060%#
62) 4-Bromofluorobenzene	11.079	95	110876	48.137	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.280%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	6104	7.037	ug/l	91
20) Methylene Chloride	2.788	84	2939	1.880	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
Data File : VX038983.D
Acq On : 21 Nov 2023 14:41
Operator : JC/MD
Sample : 05431-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-01-20231113

Quant Time: Nov 22 03:20:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

