

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
 Data File : VX038984.D
 Acq On : 21 Nov 2023 15:04
 Operator : JC/MD
 Sample : 05405-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB05-20231113

Quant Time: Nov 22 03:20:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	99444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	204936	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83846	52.326	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.660%
35) Dibromofluoromethane	5.385	113	66861	45.173	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.340%
50) Toluene-d8	8.647	98	264171	46.649	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.300%
62) 4-Bromofluorobenzene	11.079	95	113828	50.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.420%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	8340	9.777	ug/l	97
20) Methylene Chloride	2.788	84	4148	2.698	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
Data File : VX038984.D
Acq On : 21 Nov 2023 15:04
Operator : JC/MD
Sample : 05405-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB05-20231113

Quant Time: Nov 22 03:20:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

