

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112123\  
 Data File : VX038991.D  
 Acq On : 21 Nov 2023 17:44  
 Operator : JC/MD  
 Sample : 05407-06  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 WEST-DITCH-IN-20231114

Quant Time: Nov 22 03:23:32 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111623W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Nov 16 22:59:22 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)    |
|-----------------------------|--------|-------|----------|----------|-------|-------------|
| -----                       |        |       |          |          |       |             |
| Internal Standards          |        |       |          |          |       |             |
| 1) Pentafluorobenzene       | 5.549  | 168   | 97535    | 50.000   | ug/l  | 0.00        |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 202880   | 50.000   | ug/l  | 0.00        |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 213957   | 50.000   | ug/l  | 0.00        |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 110430   | 50.000   | ug/l  | 0.00        |
| System Monitoring Compounds |        |       |          |          |       |             |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 83715    | 53.266   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 106.540%    |
| 35) Dibromofluoromethane    | 5.385  | 113   | 66362    | 45.291   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 90.580%     |
| 50) Toluene-d8              | 8.646  | 98    | 263872   | 47.069   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 94.140%     |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 114616   | 51.073   | ug/l  | 0.00        |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 102.140%    |
| Target Compounds            |        |       |          |          |       |             |
| 16) Acetone                 | 2.373  | 43    | 1950     | 2.331    | ug/l  | Qvalue # 89 |
| -----                       |        |       |          |          |       |             |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112123\  
Data File : VX038991.D  
Acq On : 21 Nov 2023 17:44  
Operator : JC/MD  
Sample : 05407-06  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
WEST-DITCH-IN-20231114

Quant Time: Nov 22 03:23:32 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111623W.M  
Quant Title : SW846 8260  
QLast Update : Thu Nov 16 22:59:22 2023  
Response via : Initial Calibration

