

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX1112123\
 Data File : VX038993.D
 Acq On : 21 Nov 2023 18:30
 Operator : JC/MD
 Sample : 05405-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10-MW01S-20231113

Quant Time: Nov 22 03:24:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	94255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	196712	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100076	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80950	53.300	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.600%
35) Dibromofluoromethane	5.385	113	61974	43.622	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.240%
50) Toluene-d8	8.647	98	249474	45.896	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.800%#
62) 4-Bromofluorobenzene	11.079	95	107246	49.287	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.580%
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	3.111	73	6078	1.471	ug/l	97
20) Methylene Chloride	2.794	84	923	0.633	ug/l #	85
27) cis-1,2-Dichloroethene	4.495	96	1556	1.030	ug/l	86
44) Trichloroethene	7.135	130	1026	0.700	ug/l	85
64) Tetrachloroethene	9.269	164	5301	4.179	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
Data File : VX038993.D
Acq On : 21 Nov 2023 18:30
Operator : JC/MD
Sample : 05405-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10-MW01S-20231113

Quant Time: Nov 22 03:24:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

