



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112123\
Data File : VX038973.D
Acq On : 21 Nov 2023 08:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Title : SW846 8260
Last Update : Thu Nov 16 22:59:22 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	4682	PASS
75	95	30	60	51.5	12542	PASS
95	95	100	100	100.0	24341	PASS
96	95	5	9	6.4	1567	PASS
173	174	0.00	2	0.6	95	PASS
174	95	50	100	65.3	15895	PASS
175	174	5	9	6.8	1085	PASS
176	174	95	101	97.9	15561	PASS
177	176	5	9	5.7	890	PASS