

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
 Data File : VX043939.D
 Acq On : 21 Nov 2024 16:50
 Operator : JC/MD
 Sample : P4934-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20241119

Quant Time: Nov 22 04:01:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	112214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	209177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94994	49.356	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.720%
35) Dibromofluoromethane	5.379	113	68871	46.078	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	8.647	98	257979	50.071	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.140%
62) 4-Bromofluorobenzene	11.079	95	83319	47.517	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.040%
Target Compounds						
					Qvalue	
3) Chloromethane	1.300	50	381	0.255	ug/l #	74
16) Acetone	2.379	43	2060	2.330	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
Data File : VX043939.D
Acq On : 21 Nov 2024 16:50
Operator : JC/MD
Sample : P4934-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20241119

Quant Time: Nov 22 04:01:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

