

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112124\  
 Data File : VX043941.D  
 Acq On : 21 Nov 2024 17:36  
 Operator : JC/MD  
 Sample : P4934-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BPOW6-11-20241118

Quant Time: Nov 22 04:02:19 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X112124W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 22 03:45:52 2024  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                       |                |      |                     |        |        |          |
| Internal Standards          |                |      |                     |        |        |          |
| 1) Pentafluorobenzene       | 5.544          | 168  | 120827              | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.757          | 114  | 229101              | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055         | 117  | 190728              | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024         | 152  | 77315               | 50.000 | ug/l   | 0.00     |
|                             |                |      |                     |        |        |          |
| System Monitoring Compounds |                |      |                     |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.952          | 65   | 104075              | 50.220 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 74 - 125 |      | Recovery = 100.440% |        |        |          |
| 35) Dibromofluoromethane    | 5.385          | 113  | 76319               | 46.620 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = 93.240%  |        |        |          |
| 50) Toluene-d8              | 8.647          | 98   | 274495              | 48.643 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 86 - 113 |      | Recovery = 97.280%  |        |        |          |
| 62) 4-Bromofluorobenzene    | 11.079         | 95   | 89002               | 46.344 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 77 - 121 |      | Recovery = 92.680%  |        |        |          |
|                             |                |      |                     |        |        |          |
| Target Compounds            |                |      |                     |        |        |          |
| 3) Chloromethane            | 1.300          | 50   | 762                 | 0.473  | ug/l # | 78       |
| 16) Acetone                 | 2.386          | 43   | 1462                | 1.536  | ug/l   | 98       |
| -----                       |                |      |                     |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112124\  
Data File : VX043941.D  
Acq On : 21 Nov 2024 17:36  
Operator : JC/MD  
Sample : P4934-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 29 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BPOW6-11-20241118

Quant Time: Nov 22 04:02:19 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X112124W.M  
Quant Title : SW846 8260  
QLast Update : Fri Nov 22 03:45:52 2024  
Response via : Initial Calibration

