

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
 Data File : VX043943.D
 Acq On : 21 Nov 2024 18:23
 Operator : JC/MD
 Sample : P4934-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-9-20241119

Quant Time: Nov 22 04:03:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	110330	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	206208	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94499	49.937	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.880%		
35) Dibromofluoromethane	5.385	113	69376	47.084	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.160%		
50) Toluene-d8	8.647	98	251263	49.469	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.940%		
62) 4-Bromofluorobenzene	11.079	95	79292	45.871	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	91.740%		
Target Compounds						
3) Chloromethane	1.294	50	428	0.291	ug/l #	87
16) Acetone	2.386	43	1356	1.560	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
Data File : VX043943.D
Acq On : 21 Nov 2024 18:23
Operator : JC/MD
Sample : P4934-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW6-9-20241119

Quant Time: Nov 22 04:03:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

