

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
 Data File : VX043947.D
 Acq On : 21 Nov 2024 19:55
 Operator : JC/MD
 Sample : P4934-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-10-20241119

Quant Time: Nov 22 04:04:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	110999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211108	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95301	50.058	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.120%
35) Dibromofluoromethane	5.385	113	69690	46.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	259081	49.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	85216	48.154	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.300%
Target Compounds						
					Qvalue	
3) Chloromethane	1.301	50	528	0.357	ug/l	93
16) Acetone	2.392	43	1697	1.941	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
Data File : VX043947.D
Acq On : 21 Nov 2024 19:55
Operator : JC/MD
Sample : P4934-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW6-10-20241119

Quant Time: Nov 22 04:04:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

