

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
 Data File : VX043932.D
 Acq On : 21 Nov 2024 13:57
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX112124

Quant Time: Nov 22 03:54:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	50.537	-1.1	89	0.00
3 P	Chloromethane	50.000	49.100	1.8	89	0.00
4 C	Vinyl Chloride	50.000	50.118	-0.2#	88	0.00
5 T	Bromomethane	50.000	47.040	5.9	87	0.00
6 T	Chloroethane	50.000	40.921	18.2	73	0.00
7 T	Trichlorofluoromethane	50.000	53.407	-6.8	94	0.00
8 T	Diethyl Ether	50.000	47.345	5.3	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.762	-1.5	88	0.00
10 T	Methyl Iodide	50.000	48.614	2.8	86	0.00
11 T	Tert butyl alcohol	250.000	238.127	4.7	86	0.00
12 CM	1,1-Dichloroethene	50.000	48.192	3.6#	88	0.00
13 T	Acrolein	250.000	235.500	5.8	89	0.00
14 T	Allyl chloride	50.000	51.610	-3.2	88	0.00
15 T	Acrylonitrile	250.000	259.090	-3.6	89	0.00
16 T	Acetone	250.000	249.676	0.1	88	0.00
17 T	Carbon Disulfide	50.000	49.375	1.3	86	0.00
18 T	Methyl Acetate	50.000	50.643	-1.3	89	0.00
19 T	Methyl tert-butyl Ether	50.000	50.186	-0.4	88	0.00
20 T	Methylene Chloride	50.000	46.954	6.1	89	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.179	1.6	87	0.00
22 T	Diisopropyl ether	50.000	50.132	-0.3	88	0.00
23 T	Vinyl Acetate	250.000	263.441	-5.4	88	0.00
24 P	1,1-Dichloroethane	50.000	50.466	-0.9	88	0.00
25 T	2-Butanone	250.000	261.804	-4.7	89	0.00
26 T	2,2-Dichloropropane	50.000	52.343	-4.7	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.326	1.3	87	0.00
28 T	Bromochloromethane	50.000	46.920	6.2	101	0.00
29 T	Tetrahydrofuran	250.000	254.585	-1.8	89	0.00
30 C	Chloroform	50.000	49.891	0.2#	90	0.00
31 T	Cyclohexane	50.000	51.278	-2.6	89	0.00
32 T	1,1,1-Trichloroethane	50.000	51.320	-2.6	88	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.183	-0.4	108	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	52.442	-4.9	109	0.00
36 T	1,1-Dichloropropene	50.000	53.218	-6.4	91	0.00
37 T	Ethyl Acetate	50.000	52.656	-5.3	90	0.00
38 T	Carbon Tetrachloride	50.000	51.786	-3.6	88	0.00
39 T	Methylcyclohexane	50.000	53.489	-7.0	89	0.00
40 TM	Benzene	50.000	50.257	-0.5	88	0.00
41 T	Methacrylonitrile	50.000	53.851	-7.7	88	0.00
42 TM	1,2-Dichloroethane	50.000	50.421	-0.8	88	0.00
43 T	Isopropyl Acetate	50.000	51.469	-2.9	88	0.00
44 TM	Trichloroethene	50.000	44.554	10.9	89	0.00
45 C	1,2-Dichloropropane	50.000	47.811	4.4#	83	0.00
46 T	Dibromomethane	50.000	50.662	-1.3	88	0.00
47 T	Bromodichloromethane	50.000	52.648	-5.3	88	0.00
48 T	Methyl methacrylate	50.000	52.284	-4.6	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
 Data File : VX043932.D
 Acq On : 21 Nov 2024 13:57
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX112124

Quant Time: Nov 22 03:54:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	905.506	9.4	79	0.00
50 S	Toluene-d8	50.000	51.234	-2.5	108	0.00
51 T	4-Methyl-2-Pentanone	250.000	265.520	-6.2	89	0.00
52 CM	Toluene	50.000	51.507	-3.0#	88	0.00
53 T	t-1,3-Dichloropropene	50.000	53.077	-6.2	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.205	-4.4	87	0.00
55 T	1,1,2-Trichloroethane	50.000	50.069	-0.1	88	0.00
56 T	Ethyl methacrylate	50.000	53.118	-6.2	88	0.00
57 T	1,3-Dichloropropane	50.000	50.945	-1.9	89	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	235.230	5.9	85	0.00
59 T	2-Hexanone	250.000	273.011	-9.2	89	0.00
60 T	Dibromochloromethane	50.000	52.557	-5.1	87	0.00
61 T	1,2-Dibromoethane	50.000	52.367	-4.7	89	0.00
62 S	4-Bromofluorobenzene	50.000	53.987	-8.0	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	50.396	-0.8	89	0.00
65 PM	Chlorobenzene	50.000	48.094	3.8	86	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.515	-5.0	89	0.00
67 C	Ethyl Benzene	50.000	50.991	-2.0#	88	0.00
68 T	m/p-Xylenes	100.000	102.572	-2.6	87	0.00
69 T	o-Xylene	50.000	51.396	-2.8	86	0.00
70 T	Styrene	50.000	51.728	-3.5	86	0.00
71 P	Bromoform	50.000	47.449	5.1	84	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	96	0.00
73 T	Isopropylbenzene	50.000	49.222	1.6	89	0.00
74 T	N-amyl acetate	50.000	51.406	-2.8	88	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.356	1.3	90	0.00
76 T	1,2,3-Trichloropropane	50.000	49.973	0.1	89	0.00
77 T	Bromobenzene	50.000	48.081	3.8	87	0.00
78 T	n-propylbenzene	50.000	50.652	-1.3	88	0.00
79 T	2-Chlorotoluene	50.000	47.551	4.9	87	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.134	-0.3	89	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.756	0.5	87	0.00
82 T	4-Chlorotoluene	50.000	48.652	2.7	86	0.00
83 T	tert-Butylbenzene	50.000	50.165	-0.3	88	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.410	-0.8	89	0.00
85 T	sec-Butylbenzene	50.000	50.307	-0.6	87	0.00
86 T	p-Isopropyltoluene	50.000	51.511	-3.0	88	0.00
87 T	1,3-Dichlorobenzene	50.000	47.529	4.9	87	0.00
88 T	1,4-Dichlorobenzene	50.000	47.306	5.4	87	0.00
89 T	n-Butylbenzene	50.000	52.020	-4.0	86	0.00
90 T	Hexachloroethane	50.000	50.050	-0.1	87	0.00
91 T	1,2-Dichlorobenzene	50.000	47.656	4.7	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.030	-2.1	91	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.495	1.0	87	0.00
94 T	Hexachlorobutadiene	50.000	50.065	-0.1	92	0.00
95 T	Naphthalene	50.000	51.831	-3.7	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
Data File : VX043932.D
Acq On : 21 Nov 2024 13:57
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICV VX112124

Quant Time: Nov 22 03:54:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.246	-2.5	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6