

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048642.D
 Acq On : 21 Nov 2025 09:37
 Operator : JC/MD
 Sample : VX1121MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1121MBL01

Quant Time: Nov 21 22:07:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	125703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	249754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	268877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	153820	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	102308	58.409	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.820%
35) Dibromofluoromethane	5.367	113	90916	48.096	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.200%
50) Toluene-d8	8.629	98	314712	53.382	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.760%
62) 4-Bromofluorobenzene	11.061	95	133697	60.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	121.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048642.D
Acq On : 21 Nov 2025 09:37
Operator : JC/MD
Sample : VX1121MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1121MBL01

Quant Time: Nov 21 22:07:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

