

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
 Data File : VX048655.D
 Acq On : 21 Nov 2025 14:44
 Operator : JC/MD
 Sample : Q3702-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 21 22:14:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	160298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	322084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	341432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	184333	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	122205	54.711	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.420%
35) Dibromofluoromethane	5.367	113	105455	43.259	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.520%
50) Toluene-d8	8.628	98	357166	46.978	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.960%
62) 4-Bromofluorobenzene	11.061	95	155597	54.916	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.840%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112125\
Data File : VX048655.D
Acq On : 21 Nov 2025 14:44
Operator : JC/MD
Sample : Q3702-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 21 22:14:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.531 min Scan# 729
Delta R.T. -0.007 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:168 Resp: 160298
Ion Ratio Lower Upper
168 100
99 58.2 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 54.711 ug/l
RT: 5.934 min Scan# 795
Delta R.T. -0.012 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44

Tgt Ion: 65 Resp: 122205
Ion Ratio Lower Upper
65 100
67 51.0 0.0 106.2








#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.739 min Scan# 91
Delta R.T. -0.006 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:114 Resp: 322084
Ion Ratio Lower Upper
114 100
63 20.2 0.0 41.2
88 15.6 0.0 32.2



#35
Dibromofluoromethane
Concen: 43.259 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44

Tgt Ion:113 Resp: 105455
Ion Ratio Lower Upper
113 100
111 102.4 81.9 122.9
192 19.7 15.8 23.6









#50
Toluene-d8
Concen: 46.978 ug/l
RT: 8.628 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion: 98 Resp: 357166
Ion Ratio Lower Upper
98 100
100 66.3 53.6 80.4



#62
4-Bromofluorobenzene
Concen: 54.916 ug/l
RT: 11.061 min Scan# 1636
Delta R.T. -0.000 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44

Tgt Ion: 95 Resp: 155597
Ion Ratio Lower Upper
95 100
174 79.9 0.0 161.8
176 77.4 0.0 153.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tgt Ion:117 Resp: 341432
Ion Ratio Lower Upper
117 100
82 56.1 44.4 66.6
119 32.4 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX048655.D
Acq: 21 Nov 2025 14:44

Tgt Ion:152 Resp: 184333
Ion Ratio Lower Upper
152 100
115 58.9 39.2 117.6
150 157.7 0.0 347.0

