

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112219\
 Data File : VX013531.D
 Acq On : 22 Nov 2019 14:25
 Operator : JC/SP
 Sample : K5958-07DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5DL

Quant Time: Nov 22 15:30:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	694613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1120793	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1000461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	458439	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	410040	49.83	ug/l	0.00
Spiked Amount			Recovery	=	99.66%	
35) Dibromofluoromethane	5.49	113	337414	48.00	ug/l	0.00
Spiked Amount			Recovery	=	96.00%	
50) Toluene-d8	8.71	98	1355506	51.15	ug/l	0.00
Spiked Amount			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	11.14	95	458430	47.93	ug/l	0.00
Spiked Amount			Recovery	=	95.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
67) Ethyl Benzene	10.25	91	153861	4.401	ug/l	99
68) m/p-Xylenes	10.35	106	32776	2.479	ug/l	93
69) o-Xylene	10.70	106	26201	2.024	ug/l	96
73) Isopropylbenzene	11.01	105	144002	4.738	ug/l	99
78) n-propylbenzene	11.35	91	289725	8.557	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	393678	15.550	ug/l	100
83) tert-Butylbenzene	11.77	119	17650	0.693	ug/l	94
84) 1,2,4-Trimethylbenzene	11.81	105	1572701	61.924	ug/l	97
85) sec-Butylbenzene	11.94	105	76673	2.618	ug/l	78
86) p-Isopropyltoluene	12.06	119	38266	1.419	ug/l	96
89) n-Butylbenzene	12.38	91	58399	2.481	ug/l #	45
95) Naphthalene	13.83	128	72761	2.303	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112219\
Data File : VX013531.D
Acq On : 22 Nov 2019 14:25
Operator : JC/SP
Sample : K5958-07DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5DL

Quant Time: Nov 22 15:30:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

